

Month	Year
5	2022

Day of Month	Non Business Day	Time of reading 24-hr clock	Refrigerator 1 Logger ③			Refrigerator 2 Logger ⑤			Refrigerator 4 Logger ⑨			Refrigerator 5 Logger ⑫			Room Logger ①			MONITORING (✓)	INITIALS
			Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)		
1	✓	___:___																N.E	
2		13:10	5.1		✓	5.3	✓	5.2	✓	5.6	✓	21.6	✓	✓				N.E	
3	✓	___:___																N.E	
4	✓	___:___																N.E	
5	✓	___:___																N.E	
6		13:25	5.3		✓	5.3	✓	4.4	✓	6.0	✓	22.1	✓	✓				N.E	
7	✓	___:___																N.E	
8	✓	___:___																N.E	
9		13:10	4.7		✓	5.6	✓	4.3	✓	4.9	✓	21.8	✓	✓				N.E	
10		13:15	5.8		✓	6.0	✓	4.9	✓	5.1	✓	22.0	✓	✓				N.E	
11		14:30	4.6		✓	5.2	✓	5.4	✓	5.3	✓	21.9	✓	✓				N.E	
12		14:55	5.2		✓	5.5	✓	4.7	✓	5.6	✓	21.9	✓	✓				N.E	
13		13:15	4.9		✓	4.3	✓	4.3	✓	5.1	✓	22.0	✓	✓				N.E	
14	✓	___:___																N.E	
15	✓	___:___																N.E	
16		14:30	4.3		✓	4.5	✓	4.5	✓	5.3	✓	21.9	✓	✓				N.E	
17		14:30	4.3		✓	5.3	✓	4.2	✓	5.9	✓	21.9	✓	✓				N.E	
18		15:10	5.0		✓	4.8	✓	4.3	✓	5.0	✓	21.8	✓	✓				N.E	
19		13:25	5.8		✓	6.0	✓	5.0	✓	5.2	✓	21.8	✓	✓				N.E	
20		14:00	5.6		✓	5.2	✓	5.0	✓	6.1	✓	21.8	✓	✓				N.E	
21	✓	___:___																N.E	
22	✓	___:___																N.E	
23		13:15	4.8		✓	4.4	✓	4.6	✓	5.7	✓	22.0	✓	✓				N.E	
24		13:40	4.7		✓	4.9	✓	4.8	✓	5.0	✓	22.0	✓	✓				N.E	
25		13:40	5.1		✓	5.3	✓	4.6	✓	6.1	✓	22.0	✓	✓				N.E	
26		13:20	4.4		✓	4.9	✓	4.7	✓	6.0	✓	22.0	✓	✓				N.E	
27		13:35	4.4		✓	4.9	✓	4.9	✓	6.2	✓	21.7	✓	✓				N.E	
28	✓	___:___																N.E	
29	✓	___:___																N.E	
30		18:50	5.0		✓	5.0	✓	5.8	✓	6.2	✓	22.0	✓	✓				N.E	
31		13:45	4.8		✓	4.4	✓	4.9	✓	5.6	✓	21.8	✓	✓				N.E	

Refrigerator No.		1	2	4	5	INITIALS
Day of Month	Time of reading 24-hr clock	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	
5	18:50	✓	✓	✓	✓	N.E