

Month	Year
1	2024

Day of Month	Non Business Day	Time of reading 24-hr clock	Refrigerator 1 Logger ④			Refrigerator 2 Logger ⑥			Refrigerator ⑩			Refrigerator ⑧ Logger ③			Refrigerator ⑨ Logger ⑮			Room Logger ②			MONITORING (C)	INITIALS
			Temperature (C)	Alert (V)	NoAlert (S)	Temperature (C)	Alert (V)	NoAlert (S)	Temperature (C)	Alert (V)	NoAlert (S)	Temperature (C)	Alert (V)	NoAlert (S)	Temperature (C)	Alert (V)	NoAlert (S)	Temperature (C)	Alert (V)	NoAlert (S)		
1	✓																					
2	✓																					
3	✓																					
4		15:45	4.7	✓		4.9	✓		5.5	✓		4.2	✓				21.8	✓	✓		M.K	
5		15:25	4.6	✓		4.7	✓		5.6	✓		4.5	✓				21.8	✓	✓		M.K	
6	✓																					
7	✓																					
8	✓																					
9		15:35	4.7	✓		4.6	✓		5.6	✓		4.1	✓				21.6	✓	✓		M.K	
10		13:20	4.6	✓		4.7	✓		5.6	✓		4.4	✓				21.9	✓	✓		M.K	
11		15:35	4.8	✓		4.7	✓		5.7	✓		4.7	✓				21.9	✓	✓		M.K	
12		14:55	4.6	✓		4.8	✓		5.6	✓		4.9	✓				21.9	✓	✓		M.K	
13	✓																					
14	✓																					
15		14:35	4.6	✓		4.9	✓		5.5	✓		4.6	✓				21.9	✓	✓		M.K	
16		15:30	4.8	✓		4.9	✓		5.5	✓		4.6	✓				21.9	✓	✓		M.K	
17		15:15	4.6	✓		4.9	✓		5.3	✓		4.9	✓				21.8	✓	✓		K.S	
18		15:20	4.8	✓		4.8	✓		5.2	✓		5.0	✓				22.1	✓	✓		K.S	
19		15:20	4.6	✓		4.8	✓		5.2	✓		4.9	✓				22.1	✓	✓		K.S	
20	✓											4.3	✓									
21	✓																					
22		14:55	4.7	✓		4.9	✓		5.3	✓		4.3	✓				22.0	✓	✓		K.S	
23		15:35	4.9	✓		4.8	✓		5.2	✓		4.7	✓				21.7	✓	✓		K.S	
24		14:35	4.6	✓		4.9	✓		5.3	✓		4.3	✓				21.5	✓	✓		K.S	
25		15:10	4.8	✓		4.9	✓		5.3	✓		4.4	✓				21.7	✓	✓		K.S	
26		15:55	4.6	✓		4.9	✓		5.5	✓		4.6	✓				21.6	✓	✓		K.S	
27	✓																					
28	✓																					
29		14:45	4.7	✓		5.0	✓		5.6	✓		4.6	✓				21.7	✓	✓		K.S	
30		15:35	4.8	✓		5.0	✓		5.5	✓		4.2	✓				21.8	✓	✓		K.S	
31		15:10	4.5	✓		4.8	✓		5.5	✓		4.4	✓				21.9	✓	✓		K.S	

Refrigerator No.	1	2	4	5	8	INITIALS	
Day of Month	Time of reading 24-hr clock	No abnormality (C)	No abnormality (C)	No abnormality (C)	No abnormality (C)	No abnormality (C)	INITIALS
15	08:30	✓	✓	✓	✓	✓	K.S

7:40-
 1/26 14-④ 電池のため冷蔵庫から取り出し
 電池交換を行った。
 温度が正常に戻った(記録) してしまえ
 確認L.10201 - 再度測定再開した。 K.S