

Month	Year
12	2022

Day of Month	Non Business Day	Time of reading 24-hr clock	Refrigerator 1 Logger ③			Refrigerator 2 Logger ⑤			Refrigerator 4 Logger ⑨			Refrigerator 5 Logger ⑫			Room Logger ①			MONITORING (✓)	INITIALS
			Temperature (°C)	Alert (-/✓)	NoAlert (✓)	Temperature (°C)	Alert (-/✓)	NoAlert (✓)	Temperature (°C)	Alert (-/✓)	NoAlert (✓)	Temperature (°C)	Alert (-/✓)	NoAlert (✓)	Temperature (°C)	Alert (-/✓)	NoAlert (✓)		
1		14:45	5.6		✓	4.6		✓	5.7		✓	5.6		✓	21.5		✓	K.S	
2		14:00	4.9		✓	3.5		✓	5.2		✓	5.8		✓	21.5		✓	K.S	
3	✓	___:___																	
4	✓	___:___																	
5		14:40	4.6		✓	3.4		✓	4.7		✓	5.8		✓	21.6		✓	K.S	
6		14:20	5.2		✓	4.7		✓	5.6		✓	6.1		✓	21.3		✓	K.S	
7		14:15	4.7		✓	5.3		✓	5.8		✓	6.1		✓	21.6		✓	K.S	
8		12:55	5.6		✓	3.8		✓	5.3		✓	5.5		✓	21.5		✓	K.S	
9		15:00	4.6		✓	4.7		✓	5.7		✓	5.6		✓	21.4		✓	K.S	
10	✓	___:___																	
11	✓	___:___																	
12		13:40	5.1		✓	4.6		✓	4.9		✓	5.6		✓	21.6		✓	K.S	
13		14:20	4.8		✓	4.1		✓	5.8		✓	6.4		✓	21.8		✓	K.S	
14		14:35	5.5		✓	4.6		✓	4.9		✓	5.5		✓	21.8		✓	K.S	
15		14:55	4.5		✓	3.7		✓	5.4		✓	5.7		✓	21.8		✓	K.S	
16		14:20	4.5		✓	3.8		✓	5.6		✓	6.2		✓	21.7		✓	K.S	
17	✓	___:___																	
18	✓	___:___																	
19		14:25	5.4		✓	4.3		✓	5.3		✓	6.0		✓	21.6		✓	K.S	
20		15:00	4.4		✓	5.3		✓	5.5		✓	6.2		✓	21.6		✓	K.S	
21		15:55	5.3		✓	5.5		✓	5.9		✓	5.3		✓	21.3		✓	K.S	
22		14:35	4.4		✓	3.9		✓	4.9		✓	5.5		✓	21.3		✓	K.S	
23		14:15	6.0		✓	3.6		✓	5.1		✓	5.3		✓	21.4		✓	K.S	
24	✓	___:___																	
25	✓	___:___																	
26		14:15	5.7		✓	4.1		✓	5.9		✓	6.4		✓	21.5		✓	K.S	
27		14:10	4.8		✓	3.5		✓	5.5		✓	5.6		✓	21.5		✓	K.S	
28		13:55	5.2		✓	4.2		✓	4.7		✓	6.0		✓	21.3		✓	K.S	
29	✓	___:___																	
30	✓	___:___																	
31	✓	___:___																	

Refrigerator No.		1	2	4	5	INITIALS
Day of Month	Time of reading 24-hr clock	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	
22	17:20	✓	✓	✓	✓	K.S