

Month	Year
10	2023

Day of Month	Non Business Day	Time of reading 24-hr clock	Refrigerator 1 Logger ④			Refrigerator 2 Logger ⑥			Refrigerator 4 Logger ⑩			Refrigerator 5 Logger ⑧			Refrigerator 6 Logger ⑮			Room Logger ②			MONITORING (✓)	INITIALS
			Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)		
1	✓	___:___																				
2		15:15	4.9		✓				5.3		✓	4.5		✓	4.8		✓	23.0		✓	✓	K.S
3		15:45	4.7		✓	5.1		✓	5.3		✓	4.1		✓	4.3		✓	22.8		✓	✓	K.S
4		15:50	4.9		✓	5.3		✓	5.4		✓	4.2		✓	4.6		✓	22.8		✓	✓	K.S
5		15:45	4.8		✓	5.0		✓	5.5		✓	4.2		✓	3.5		✓	22.8		✓	✓	K.S
6		15:00	5.0		✓	4.9		✓	5.5		✓	4.3		✓	3.3		✓	22.9		✓	✓	K.S
7	✓	___:___																				
8	✓	___:___																				
9	✓	___:___																				
10		15:05	4.9		✓	4.7		✓	5.6		✓	4.3		✓	4.0		✓	22.6		✓	✓	K.S
11		15:30	4.9		✓	4.8		✓	5.5		✓	4.1		✓	3.7		✓	22.6		✓	✓	K.S
12		15:40	4.9		✓	5.0		✓	5.5		✓	4.4		✓	3.9		✓	22.8		✓	✓	K.S
13		15:50	5.0		✓	5.0		✓	5.7		✓	4.0		✓	3.7		✓	22.7		✓	✓	K.S
14	✓	___:___																				
15	✓	___:___																				
16		15:00	4.8		✓	5.0		✓	5.7		✓	4.0		✓	3.7		✓	22.3		✓	✓	K.S
17		15:30	4.9		✓	5.2		✓	5.6		✓	3.9		✓				22.7		✓	✓	K.S
18		15:20	4.9		✓	5.1		✓	5.5		✓	4.2		✓				22.5		✓	✓	K.S
19		15:40	4.9		✓	5.0		✓	5.5		✓	3.8		✓				22.8		✓	✓	K.S
20		15:10	5.1		✓	5.1		✓	5.6		✓	4.4		✓				22.9		✓	✓	K.S
21	✓	___:___																				
22	✓	___:___																				
23		15:50	5.0		✓	5.1		✓	5.5		✓	4.2		✓				22.5		✓	✓	K.S
24		14:50	4.8		✓	5.1		✓	5.5		✓	4.2		✓				22.5		✓	✓	K.S
25		09:20	4.9		✓	5.0		✓	5.5		✓	4.1		✓				22.1		✓	✓	K.S
26		15:10	5.0		✓	5.0		✓	5.5		✓	3.8		✓				22.7		✓	✓	K.S
27		15:30	4.9		✓	5.0		✓	5.6		✓	4.6		✓				22.7		✓	✓	K.S
28	✓	___:___																				
29	✓	___:___																				
30		15:00	4.9		✓	5.0		✓	5.7		✓	4.7		✓				22.3		✓	✓	K.S
31		14:26	4.9		✓	5.0		✓	5.6		✓	4.5		✓				22.6		✓	✓	K.S

Refrigerator No.		1	2	4	5	6	INITIALS
Day of Month	Time of reading 24-hr clock	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	
25	09:20	✓	✓	✓	✓	✓	K.S

⑥ 10/3 8:00- 温度管理再開

⑮
⑭ 電池残量少いため修理

10/17 15時冷蔵庫から取り出して