

Month	Year
11	2023

Day of Month	Non Business Day	Time of reading 24-hr clock	Refrigerator 1 Logger ③			Refrigerator 2 Logger ⑤			Refrigerator 4 Logger ⑨			Refrigerator 5 Logger ⑫			Refrigerator 6 Logger ⑮			Room Logger ①			MONITORING (✓)	INITIALS
			Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)	Temperature (°C)	Alert (✓)	NoAlert (✓)		
1		09:30	4.5		✓	3.7		✓									22.0		✓	✓	M.K	
2		09:50	5.2		✓	4.5		✓									22.1		✓	✓	M.K	
3	✓	___:___																				
4	✓	___:___																				
5	✓	___:___																				
6		15:25	4.9		✓	4.0		✓				6.2		✓			22.4		✓	✓	M.K	
7		15:45	5.4		✓	3.1		✓				5.3		✓			22.0		✓	✓	M.K	
8		15:25	3.0		✓	2.6		✓				5.3		✓			22.1		✓	✓	M.K	
9		15:27	2.9		✓	3.7		✓				5.5		✓			22.3		✓	✓	K.S	
10		15:40	5.5		✓	2.8		✓				5.3		✓			22.2		✓	✓	K.S	
11	✓	___:___																				
12	✓	___:___																				
13		15:30	4.0		✓	4.6		✓				5.8		✓			21.7		✓	✓	K.S	
14		15:20	4.1		✓	2.4		✓				5.8		✓			22.0		✓	✓	K.S	
15		15:15	5.3		✓	4.3		✓				5.3		✓			21.8		✓	✓	K.S	
16		15:30	3.6		✓	4.3		✓				5.6		✓			22.1		✓	✓	K.S	
17		15:25	4.3		✓	3.9		✓				4.8		✓			21.8		✓	✓	K.S	
18	✓	___:___																				
19	✓	___:___																				
20		15:00	3.7		✓	4.1		✓				4.9		✓			21.6		✓	✓	K.S	
21		15:20	3.9		✓	2.6		✓				5.2		✓			21.9		✓	✓	K.S	
22		15:40	2.5		✓							5.5		✓			21.9		✓	✓	K.S	
23	✓	15:35 *																				
24		15:35	3.7		✓							5.4		✓			21.7		✓	✓	K.S	
25	✓	___:___																				
26	✓	___:___																				
27		15:05	5.6		✓							4.7		✓			22.0		✓	✓	K.S	
28		14:30	4.1		✓							4.8		✓			21.9		✓	✓	K.S	
29		15:20	3.3		✓							4.9		✓			22.0		✓	✓	K.S	
30		15:55	2.1		✓							5.3		✓			21.8		✓	✓	K.S	
31		___:___																				

Refrigerator No.	1	2	4	5	6	INITIALS	
Day of Month	Time of reading 24-hr clock	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	INITIALS
1	09:30	✓	✓	✓	✓	✓	K.S

11/22 15:00
 口が⑤ 電池残量少のため回収
 冷蔵庫から取り出した

口が⑫
 11/9 8:00-記録再開

*error 23/Nov/2023 K.S