

Month	Year
2	2024

Day of Month	Non Business Day	Time of reading 24-hr clock	Refrigerator 1 Logger ③			Refrigerator 2 Logger ⑤			Refrigerator 4 Logger ⑨			Refrigerator 5 Logger ⑫			Refrigerator 6 Logger ⑮			Room Logger ①			MONITORING (✓)	INITIALS
			Temperatur (°C)	Alert (✓)	NoAlert (✓)	Temperatur (°C)	Alert (✓)	NoAlert (✓)	Temperatur (°C)	Alert (✓)	NoAlert (✓)	Temperatur (°C)	Alert (✓)	NoAlert (✓)	Temperatur (°C)	Alert (✓)	NoAlert (✓)	Temperatur (°C)	Alert (✓)	NoAlert (✓)		
1		15:50	4.5		✓	4.2		✓	4.5		✓	2.8		✓			21.9		✓	✓	K.S	
2		16:20	4.3		✓							3.0		✓						✓	K.S	
3	✓	---																				
4	✓	---																				
5		16:30	4.3		✓							3.1		✓						✓	K.S	
6		14:50	3.8		✓							3.2		✓						✓	K.S	
7		15:50	4.9		✓							2.7		✓						✓	K.S	
8		15:45	3.9		✓							2.6		✓						✓	K.S	
9		16:00	5.1		✓							3.3		✓						✓	K.S	
10	✓	---																				
11	✓	---																				
12	✓	---																				
13		15:40	4.0		✓							3.0		✓						✓	K.S	
14		15:35	4.9		✓							3.0		✓						✓	K.S	
15		16:50	4.7		✓							3.4		✓						✓	K.S	
16		15:10	4.9		✓							5.1		✓						✓	K.S	
17	✓	---																				
18	✓	---																				
19		16:30	3.6		✓							5.2		✓						✓	K.S	
20		16:45	5.5		✓							3.6		✓						✓	K.S	
21		15:50	5.4		✓							4.6		✓						✓	K.S	
22		16:15				3.3		✓	4.4		✓						21.7		✓	✓	K.S	
23	✓	---																				
24	✓	---																				
25	✓	---																				
26		15:45				3.7		✓	4.4		✓						21.8		✓	✓	K.S	
27		15:15				3.9		✓	3.3		✓						21.7		✓	✓	K.S	
28		15:45				4.1		✓	4.1		✓						21.8		✓	✓	K.S	
29		15:55				4.2		✓	3.9		✓						21.9		✓	✓	K.S	
30		---																				
31		---																				

Refrigerator No.		1	2	4	5	6	INITIALS
Day of Month	Time of reading 24-hr clock	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	
22	09:00	✓	✓	✓	✓	✓	K.S

7/16:00-
校正のため下記ロガーを管理庫から出した。
ロガー-⑤⑨①

7/22 12:45-
校正から戻ってきた下記ロガーの温度記録を再開した
ロガー-⑤⑨①

7/22 13:50-
校正のため、下記ロガーを管理庫から出した
ロガー-③⑬⑭