

01 Jul 2026 Poshi Salt

Storage Location: Pharmacy of Nagasaki University Hospital

Month	Year
6	2026

6		2026		Refrigerator 1			Refrigerator 2			Refrigerator 4			Refrigerator 5			Refrigerator 6			Room			MONITORING (S)	INITIALS
Logger ④			Logger ⑥			Logger ⑩			Logger ⑧			Logger ⑮			Logger ②								
Day of Month	Non Business Day	Time of reading 24-hr clock	Temperatur ° (C)	Alert (✓)	NoAlert (✓)	Temperatur ° (C)	Alert (✓)	NoAlert (✓)	Temperatur ° (C)	Alert (✓)	NoAlert (✓)	Temperatur ° (C)	Alert (✓)	NoAlert (✓)	Temperatur ° (C)	Alert (✓)	NoAlert (✓)	Temperatur ° (C)	Alert (✓)	NoAlert (✓)			
1		16:30	5.1		✓	4.6		✓	5.8		✓	5.4		✓	4.6		✓	21.8		✓	✓	S.K	
2		16:50	5.1		✓	4.5		✓	5.8		✓	5.5		✓	4.9		✓	21.5		✓	✓	S.K	
3		16:40	4.9		✓	4.6		✓	5.7		✓	5.5		✓	5.3		✓	21.0		✓	✓	S.K	
4		16:50	5.1		✓	4.6		✓	5.8		✓	5.4		✓	5.2		✓	21.6		✓	✓	K.W	
5		14:20	4.9		✓	4.6		✓	5.7		✓	5.4		✓	5.1		✓	21.3		✓	✓	S.K	
6	✓	_____:																					
7	✓	_____:																					
8		16:45	5.0		✓	4.5		✓	5.8		✓	5.5		✓	5.1		✓	21.4		✓	✓	M.K	
9		16:55	4.9		✓	4.4		✓	5.7		✓	5.5		✓	4.6		✓	21.4		✓	✓	S.K	
10		16:30	5.2		✓	4.3		✓	5.7		✓	5.4		✓	4.7		✓	21.7		✓	✓	S.K	
11		16:10	5.0		✓	4.9		✓	5.3		✓	5.4		✓	5.3		✓	21.8		✓	✓	S.F	
12		16:05	5.2		✓	5.0		✓	5.7		✓	5.6		✓	5.1		✓	21.6		✓	✓	T.S	
13	✓	_____:																					
14	✓	_____:																					
15		16:45	5.2		✓	5.0		✓	5.7		✓	5.6		✓	5.3		✓	21.7		✓	✓	S.K	
16		17:00	5.0		✓	5.0		✓	5.7		✓	5.6		✓	4.7		✓	21.7		✓	✓	S.K	
17		16:20	5.0		✓	5.0		✓	5.7		✓	5.8		✓	4.8		✓	21.8		✓	✓	S.K	
18		16:40	5.0		✓	4.9		✓	5.5		✓	5.6		✓	4.8		✓	21.5		✓	✓	S.K	
19		16:14	5.1		✓	4.9		✓	5.6		✓	5.6		✓	5.2		✓	21.5		✓	✓	M.K	
20	✓	_____:																					
21	✓	_____:																					
22		16:30	5.1		✓	5.0		✓	5.5		✓	5.6		✓	5.4		✓	21.6		✓	✓	K.W	
23		16:50	5.1		✓	5.1		✓	5.6		✓	5.8		✓	5.0		✓	21.9		✓	✓	S.K	
24		16:30	5.2		✓				5.5		✓	5.8		✓	5.3		✓	22.2		✓	✓	S.K	
25		16:30	5.1		✓				5.2		✓	5.6		✓	4.9		✓	21.9		✓	✓	S.F	
26		16:10	5.3		✓				5.3		✓	5.7		✓	4.9		✓	21.3		✓	✓	T.S	
27	✓	_____:																					
28	✓	_____:																					
29		16:30	5.2		✓				5.2		✓	5.6		✓	5.1		✓	21.8		✓	✓	T.S	
30		16:30	4.9		✓				5.3		✓	5.7		✓	4.6		✓	21.4		✓	✓	M.K	
31		_____:																					

* No records due to the public holiday.

Refrigerator No.		1	2	4	5	6	INITIALS
Day of Month	Time of reading 24-hr clock	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	
26	16:10	✓	✗	✓	✓	✓	T.S

※故障のため
修理依頼中. 26 Jun 2026 T.S

口ガ-⑥

冷蔵庫故障のため.

6/23 18:30 モニタリング停止
T.S